

Mouse IL-10 ELISA Kit

EM2IL10 EM2IL102 EM2IL105

1346.4

Number	Description
EM2IL10	Mouse Interleukin-10 (IL-10) ELISA Kit , sufficient reagents for 96 determinations Kit Contents: Anti-Mouse IL-10 96-Well Strip Plate , 1 each Lyophilized Recombinant Mouse IL-10 Standard , 2 vials Assay Buffer , 12 ml Biotinylated Antibody Reagent , 8 ml 30X Wash Buffer , 50 ml Streptavidin-HRP Concentrate , 75 μ l Streptavidin-HRP Dilution Buffer , 14 ml TMB Substrate , 13 ml Stop Solution , 13 ml, contains 0.16 M sulfuric acid Adhesive Plate Covers , 6 each
EM2IL102	Mouse Interleukin-10 (IL-10) ELISA Kit , sufficient reagents for 2 \times 96 determinations Kit Contents: Anti-Mouse IL-10 96-Well Strip Plate , 2 each Lyophilized Recombinant Mouse IL-10 Standard , 4 vials Assay Buffer , 2 \times 12 ml Biotinylated Antibody Reagent , 2 \times 8 ml 30X Wash Buffer , 2 \times 50 ml Streptavidin-HRP Concentrate , 2 \times 75 μ l Streptavidin-HRP Dilution Buffer , 2 \times 14 ml TMB Substrate , 2 \times 13 ml Stop Solution , 2 \times 13 ml, contains 0.16 M sulfuric acid Adhesive Plate Covers , 12 each
EM2IL105	Mouse Interleukin-10 (IL-10) ELISA Kit , sufficient reagents for 5 \times 96 determinations Kit Contents: Anti-Mouse IL-10 96-Well Strip Plate , 5 each Lyophilized Recombinant Mouse IL-10 Standard , 5 vials Assay Buffer , 100 ml Biotinylated Antibody Reagent , 30 ml 30X Wash Buffer , 200 ml Streptavidin-HRP Concentrate , 250 μ l Streptavidin-HRP Dilution Buffer , 70 ml TMB Substrate , 5 \times 13 ml Stop Solution , 55 ml, contains 0.16 M sulfuric acid Adhesive Plate Covers , 30 each

For research use only – not for use in diagnostic procedures.

Storage: For maximum stability, store in a non-defrosting -20°C freezer and refer to the expiration date for frozen storage on the label. Alternatively, store at 2-8°C and refer to the expiration date for refrigerated storage. Once thawed, store at 4°C until the expiration date for refrigerated storage. Kit is shipped on dry ice.

Table of Contents

Introduction	2
Procedure Summary.....	2
Additional Materials Required.....	3
Precautions.....	3
Sample Preparation.....	4
Reagent Preparation.....	4
Assay Procedure	5
Performance Characteristics	7
Reference.....	8
Data Templates.....	9

Introduction

The Mouse Interleukin-10 (IL-10) ELISA is an *in vitro* enzyme-linked immunosorbent assay for the quantitative measurement of mouse IL-10 in serum and culture supernatants.


Procedure Summary



Step 1. Add 50 µl of Assay Buffer to each well.



Step 2. Add 50 µl of Standards or samples to each well in duplicate. Cover plate and incubate at room temperature (20-25°C) for 3 hours.



Step 3. Wash plate THREE times.




Step 4. Add 50 µl of Premixed Biotinylated Antibody Reagent to each well. Cover plate and incubate at room temperature for 1 hour.




Step 5. Wash plate THREE times.



Step 6. Dilute Streptavidin-HRP Concentrate in Dilution Buffer. Add 100 µl of this solution to each well.



Step 7. Cover and incubate plate at room temperature for 30 minutes.



Step 8. Wash plate THREE times.



Step 9. Add 100 µl of TMB Substrate to each well.



Step 10. Develop plate at room temperature for 30 minutes.



Step 11. Stop reaction by adding 100 µl of Stop Solution to each well.



Step 12. Measure the absorbance on a plate reader set at 450 and 550 nm and calculate results.

Additional Materials Required

- Precision pipettors with disposable plastic tips to deliver 5-1,000 μ l
- Plastic pipettes to deliver 5-15 ml
- A glass or plastic two-liter container to prepare Wash Buffer
- A squirt wash bottle or an automated plate washer
- 1.5 ml polypropylene or polyethylene tubes to prepare standards – do not use polystyrene, polycarbonate or glass tubes
- Disposable reagent reservoirs
- 15 ml plastic tube to prepare Streptavidin-HRP Solution
- A microcentrifuge to prepare Streptavidin-HRP Solution
- A standard ELISA reader for measuring absorbance at 450 nm and 550 nm. If a 550 nm filter is not available, the absorbance can be measured at 450 nm only. Refer to the instruction manual supplied with the instrument being used.
- Graph paper or a computerized curve-fitting statistical software package

Precautions

- All samples and reagents must be at room temperature (20-25°C) before use in the assay.
- Review these instructions carefully and verify all components against the Kit Contents list (page 1) before beginning.
- Do not use a water bath to thaw samples. Thaw samples at room temperature.
- When assaying culture medium, prepare the standard curve and sample dilutions using the same medium used to culture cells. For example, if RPMI with 10% fetal calf serum (FCS) was used to culture cells, then use RPMI with 10% FCS to dilute standard and samples. For best results, use a culture medium that contains a carrier protein such as FCS. Lack of a carrier protein in the media or addition of other compounds may compromise assay results.
- If using a multichannel pipettor, always use a new disposable reagent reservoir.
- Use new disposable pipette tips for each transfer to avoid cross-contamination.
- Use a new adhesive plate cover for each incubation step.
- Once reagents have been added to the plate, take care NOT to let plate DRY at any time during the assay.
- Avoid microbial contamination of reagents.
- Vigorous plate washing is essential.
- Avoid exposing reagents to excessive heat or light during storage and incubation.
- Do not mix reagents from different kit lots. Discard unused components after assay completion.
- Do not use glass pipettes to measure TMB Substrate. Take care not to contaminate the solution. If the solution is blue before use, DO NOT USE IT.
- Individual components may contain antibiotics and preservatives. Wear gloves while performing the assay to avoid contact with samples and reagents. Please follow proper disposal procedures.

Additional Precautions for the 5-plate Kit

- Dispense only the reagent volumes required for the number of plates being used. Do not combine leftover reagents with those reserved for additional plates.
- Use only one bottle of the TMB Substrate per 96-well plate. Do not combine leftover substrate with that reserved for other plates.
- Equilibrate to room temperature only the reagent volumes required for the number of plates being used.
- Use only one vial of Standard per 96-well plate.

Sample Preparation

- Serum and culture supernatants may be tested in this assay.
- 50 µl per well of serum, plasma or culture supernatant are required.
- Store samples to be assayed within 24 hours at 2-8°C. For long-term storage, aliquot and freeze samples at -70°C.
- Avoid repeated freeze-thaw cycles when storing samples.
- Test samples and standards must be assayed in duplicate each time the ELISA is performed.
- Gradually equilibrate samples to room temperature before beginning assay. Do not use heated water baths to thaw or warm samples.
- Mix samples by gently inverting tubes.
- If samples are clotted, grossly hemolyzed, lipemic or microbially contaminated, or if there is any question about the integrity of a sample, make a note on the template and interpret results with caution.
- If the mouse IL-10 concentration possibly exceeds the highest point of the standard curve (i.e., 3,000 pg/ml), prepare one or more 10-fold dilutions of the test sample. When testing **culture supernatants**, prepare the serial dilutions using culture medium. When testing **serum**, prepare the serial dilution using the Assay Buffer provided. For example, a 10-fold dilution is prepared by adding 50 µl of test sample to 450 µl of appropriate diluent. Mix thoroughly between dilutions before assaying.

Reagent Preparation

For procedural differences when using partial plates, look for **(PP)** throughout these instructions.

Wash Buffer

- **(PP)** When using partial plates, store the reconstituted Wash Buffer at 2-8°C.
 - **Note:** Wash buffer must be at room temperature before use in the assay. Do not use Wash Buffer if it becomes visibly contaminated during storage.
1. Label a clean glass or plastic two-liter container "Wash Buffer." The 30X Wash Buffer may have a cloudy appearance.
 2. If using a 5-plate kit, add 30 ml Wash Buffer to 870 ml water for each plate used. Otherwise, add the entire contents of the 30X Wash Buffer (50 ml) bottle to the two-liter container and dilute to a final volume of 1.5 liters with ultrapure water. Mix thoroughly.

Standards


- **(PP)** Reconstitute and use one vial of the lyophilized Standard per partial plate.
- Prepare Standards just before use and use within one hour of reconstitution. Do not store reconstituted standards.
- When testing **culture supernatant samples**, reconstitute standard with ultrapure water. Reconstitution volume is stated on the standard vial label. The standard will dissolve in approximately 1 minute. Mix by gently inverting vial. Use the sample culture medium to prepare Standard Curve dilutions.

When testing **serum samples**, reconstitute standard with ultrapure water. Reconstitution volume is stated on the standard vial label. The standard will dissolve in approximately 1 minute. Mix by gently inverting vial. Use the Assay Buffer provided to prepare standard curve serial dilutions.

When testing **both serum and cell culture supernatant samples** on the same plate, validate the media to establish if the same standard curve can be used for both sample types. Prepare a standard curve (including a zero/blank) using culture medium to reconstitute and dilute the standard. Perform this curve in parallel with a standard curve prepared with Assay Buffer. If the OD values of the two curves are within 10% of the mean for both curves, then the assay can be performed with Assay Buffer, whether testing culture supernatant or serum samples.

1. Label six tubes, one for each standard curve point: 3,000 pg/ml, 1,000 pg/ml, 333 pg/ml, 111 pg/ml, 37 pg/ml and 0 pg/ml, then prepare 1:3 serial dilutions for the standard curve as follows:
2. Pipette 400 µl of appropriate diluent into each tube.

- Pipette 200 μ l of the reconstituted standard into the first tube (i.e., 3,000 pg/ml) and mix.
- Pipette 200 μ l of this dilution into the second tube (i.e., 1,000 pg/ml) and mix.
- Repeat serial dilutions four more times.



Assay Procedure

A. Sample Incubation

- (PP)** Determine number of strips required. Leave these strips in the plate frame. Tightly seal the remaining unused strips in the provided foil pouch with desiccant and store at 2-8°C. After completing assay, retain plate frame for second partial plate. When using the second partial plate, place strips securely in the plate frame.
 - Use the Data Template provided to record the locations of the zero standard (blank or negative control), IL-10 standards and test samples. Perform five standard points and one blank in duplicate with each series of unknown samples.
 - If using a multichannel pipettor, use a new reagent reservoir to add the Assay Buffer. Remove from the vial only the amount required for the number of strips being used.
- Add 50 μ l of Assay Buffer to all wells being used.
 - Add 50 μ l of reconstituted standards or test samples in duplicate to each well.
 - Note:** If the mouse IL-10 concentration in any test sample possibly exceeds the highest point on the standard curve, 3,000 pg/ml, dilute the sample as indicated in the Sample Preparation Section.
 - Carefully cover plate with an adhesive plate cover. Ensure all edges and strips are tightly sealed by running your thumb over edges and down each strip. Incubate for 3 hours at room temperature, 20-25°C.
 - Carefully remove adhesive plate cover. Wash plate THREE times with Wash Buffer as described in the Plate Washing section (section B).

B. Plate Washing

- Gently squeeze the long sides of the plate frame before washing to ensure all strips securely remain in the frame.
- Empty plate contents. Use a squirt bottle to vigorously fill each well completely with Wash Buffer, then empty plate contents. Repeat procedure two times for a total of THREE washes. Blot plate onto paper towels or other absorbent material.

Note: For automated washing, aspirate all wells and wash THREE times with Wash Buffer, overfilling wells with Wash Buffer. Blot plate onto paper towels or other absorbent material.

C. Biotinylated Antibody Reagent Incubation

Note: If using a multichannel pipettor, use a new reagent reservoir to add the Biotinylated Antibody Reagent.

(PP) Remove from the vial only the amount required for the number of strips being used.

- Add 50 μ l of the Biotinylated Antibody Reagent to each well.
- Carefully cover the plate with an adhesive plate cover. Ensure all edges and strips are tightly sealed by running your thumb over the edges and down each strip. Incubate for one (1) hour at room temperature, 20-25°C.
- Carefully remove the adhesive plate cover. Wash the plate THREE times with Wash Buffer as described in the Plate Washing section (section B).

D. Streptavidin-HRP Solution Preparation and Incubation

- Prepare Streptavidin-HRP Solution immediately before use. Do not prepare more solution than required.
 - Do not store prepared Streptavidin-HRP Solution.
 - Use a 15 ml plastic tube to prepare Streptavidin-HRP Solution.
 - If using a multichannel pipettor, use new reagent reservoir and pipette tips when adding the prepared Streptavidin-HRP Solution.
1. Centrifuge Streptavidin-HRP Concentrate to force entire vial contents to the bottom.
 2. For one complete 96-well plate, add 30 μ l of Streptavidin-HRP Concentrate to 12 ml of Streptavidin-HRP Dilution Buffer and mix gently.
 3. (PP) Use only the Streptavidin-HRP Solution amount required for the number of strips being used. For each strip, mix 2.5 μ l of Streptavidin-HRP Concentrate with 1 ml of Streptavidin-HRP Dilution Buffer. Store Streptavidin-HRP Concentrate reserved for additional strips at 2-8°C.
 4. Add 100 μ l of prepared Streptavidin-HRP Solution to each well.
 5. Carefully attach a new adhesive plate cover, ensuring all edges and strips are tightly sealed. Incubate the plate for 30 minutes at room temperature, 20-25°C.
 6. Carefully remove the adhesive plate cover, discard plate contents and wash THREE times as described in the Plate Washing section (section B).

E. Substrate Incubation and Stop Step

- Use new disposable reagent reservoirs when adding the TMB Substrate and Stop Solution.
 - Dispense from bottle ONLY amount required, 100 μ l per well, for the number of wells being used. Do not use a glass pipette to measure the TMB Substrate.
 - (PP) Do not combine leftover substrate with that reserved for the second partial plate. Take care not to contaminate remaining TMB Substrate.
1. Pipette 100 μ l of TMB Substrate Solution into each well.
 2. Allow color to develop at room temperature in the dark for 30 minutes. Do not cover plate with aluminum foil or a plate sealer. The substrate reaction yields a blue solution that turns yellow when Stop Solution is added.
 3. After 30 minutes, stop the reaction by adding 100 μ l of Stop Solution to each well.


F. Absorbance Measurement

Note: Evaluate the plate within 30 minutes of stopping the reaction.

Measure absorbance on an ELISA plate reader set at 450 nm and 550 nm. Subtract 550 nm values from 450 nm values to correct for optical imperfections in the microplate. If an absorbance at 550 nm is not available, measure the absorbance at 450 nm only. When the 550 nm measurement is omitted, absorbance values will be higher.

G. Calculation of Results

- The standard curve is used to determine IL-10 amount in an unknown sample. Generate the standard curve by plotting the average absorbance obtained for each Standard concentration on the vertical (Y) axis vs. the corresponding IL-10 concentration (pg/ml) on the horizontal (X) axis.
- Calculate results using graph paper or curve-fitting statistical software. Determine the IL-10 amount in each sample by interpolating from the absorbance value (Y axis) to concentration (X axis) using the standard curve.



- If the test sample was diluted, multiply the interpolated value obtained from the standard curve by the dilution factor to calculate pg/ml of IL-10 in the sample.
- Absorbance values obtained for duplicates should be within 10% of the mean value. Carefully consider duplicate values that differ from the mean by greater than 10%.

Performance Characteristics

Sensitivity: < 12 pg/ml of mouse IL-10

The sensitivity or lower limit of detection (LLD)¹ was determined by assaying replicates of zero and the standard curve. The mean signal of zero + 2 standard deviations read in dose from the standard curve is the LLD. This value is the smallest dose that is not zero with 95% confidence.

Assay Range: 37-3,000 pg/ml

Suggested standard curve points are 3,000; 1,000; 333; 111; 37 and 0 pg/ml.

Reproducibility:

Intra-assay CV: < 10%

Inter-assay CV: < 10%

Specificity

The following cytokines do not interfere with or cross-react in the mouse IL-10 ELISA: mouse IL-1 α , IL-1 β , IL-2, IL-3, IL-4, IL-5, IL-6, IL-7, GM-CSF, IFN γ , TNF α or human IL-10.

Recovery

Recovery in this ELISA was determined by spiking three different levels of recombinant mouse IL-10 into eight serum samples collected from apparently healthy Balb/c and NSA mice. Recoveries are as follows:

<u>Spike Level</u>	<u>Mean Recovery</u>
100 pg/ml	116%
500 pg/ml	102%
1,500 pg/ml	101%

Expected Values*


The average level of mouse IL-10 detected in eight serum samples from apparently healthy Balb/c mice was 18 pg/ml ranging from 0 to 53 pg/ml. To induce mouse IL-10 production Balb/c mice were injected with LPS, 50 μ g/mouse. Serum samples were harvested by heart puncture during a three-hour time course and were then evaluated with the Mouse IL-10 ELISA. Additionally, splenocytes harvested from Balb/c mice were stimulated with 20 μ g/ml LPS during a three-day time course. Supernatants were collected at various time points and evaluated with the Mouse IL-10 ELISA. The results of these two experiments are reported in the table below.

<u><i>In Vivo</i> LPS-Stimulated Serum Samples</u>		<u>LPS-Stimulated Splenocyte Supernatants</u>	
<u>Time After Injection (min)</u>	<u>Mouse IL-10 (pg/ml)</u>	<u>Days After Stimulation</u>	<u>Mouse IL-10 (pg/ml)</u>
0	30	1	79
60	1,170	2	277
90	4,225	3	852
120	1,623		
150	1,064	Non-stimulated Control	
180	497	3	0

***Note:** Strain-to-strain differences in the base levels of mouse IL-10 and nonspecific activity (up to 45 pg/ml) were observed when testing commercially available normal serum samples in this assay. For best results, test appropriate control mouse samples in every experiment.

Linearity of Dilution

Dilution linearity was determined by serially diluting 10 different positive samples. The dilutions were evaluated in the ELISA and “found” doses are plotted against the “expected” doses. An “r” value and slope of the regression line close to 1 indicate that the samples dilute linearly.



Reference

1. *Immunoassay: A Practical Guide*, Chan and Perlstein, Eds., 1987, Academic Press: New York, p71.

This product (“Product”) is warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts (“Documentation”) and to be free from defects in material and workmanship. Unless otherwise expressly authorized in writing, Products are supplied for research use only. No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than the original purchaser of the Product (“Buyer”).

No other warranties, express or implied, are granted, including without limitation, implied warranties of merchantability, fitness for any particular purpose, or non infringement. Buyer’s exclusive remedy for non-conforming Products during the warranty period is limited to replacement of or refund for the non-conforming Product(s).

There is no obligation to replace Products as the result of (i) accident, disaster or event of force majeure, (ii) misuse, fault or negligence of or by Buyer, (iii) use of the Products in a manner for which they were not designed, or (iv) improper storage and handling of the Products.

Current versions of product instructions are available at www.thermo.com/pierce. For a faxed copy, call 800-874-3723 or contact your local distributor.

© 2009 Thermo Fisher Scientific Inc. All rights reserved. Unless otherwise indicated, all trademarks are property of Thermo Fisher Scientific Inc. and its subsidiaries. Printed in the USA.

Data Templates

	1	2	3	4	5	6	7	8	9	10	11	12
A												
B												
C												
D												
E												
F												
G												
H												

	1	2	3	4	5	6	7	8	9	10	11	12
A												
B												
C												
D												
E												
F												
G												
H												